

EU- DECLARATION OF CONFORMITY


Transus Instruments BV
 Bloesemlaan 4
 3897LN Zeewolde
 The Netherlands



Hereby declares as manufacturer that the product(s):

Product type	Ultrasonic Flowmeter	Temperature (ambient)	-40°C to +60°C
Function	Gas flow measurement	Temperature (process)	per nameplate
Model	UIM Series Flowmeter	Pressure range (process)	per nameplate
		Year of manufacture	2020

Is in conformity with the requirements of the following directives and conformity was checked in accordance with the following standards:

Directive	Harmonized Standards and specifications	Remarks
Electromagnetic Compatibility (EMC) (for emission purposes): EMC Directive 2014/30/EC	61000-6-3 (2007) + A1 (2011).	Class B
EMC (for immunity purposes): MID Directive 2014/32/EC, Directive 2014/30/EC	OIML R137-1&2, Edition 2012.	Class A
"ATmosphere EXplosible" (ATEX) Directive 2014/34/EU Notified Body for EU-Type Examination: DEKRA Certification BV (0344) Meander 1051, Arnhem, The Netherlands	EN 60079-0 : 2012 + A11 : 2013 EN 60079-11 : 2012 EU-Type examination certificate DEKRA 16ATEX0014 X	Markings: UIM Series Flowmeter  II 1 G Ex ia IIC T4 Ga

The product must be installed in accordance with the applicable regulations and manufacturer's instructions.

Signed for Transus Instruments BV, Zeewolde, the Netherlands on Thursday, 05 March 2020 by

Representative: Arjan Stehouwer
 Function: General Manager

Signature: